

# Use of Electrical or Magnetic Stimulation for Generating Hip Flexion Torque

**Authors:** Tomotaka Ito (RPT, MS)<sup>1)</sup>, Akio Tsubahara (MD, DMSc)<sup>2)</sup>, Susumu Watanabe (RPT, PhD)<sup>1)</sup>

**Affiliations:** 1) Department of Rehabilitation, Faculty of Health Science and Technology,

Kawasaki University of Medical Welfare, 2) Department of Rehabilitation Medicine, Kawasaki Medical

School

## **Correspondence:**

Tomotaka Ito

288, Matsushima, Kurashiki, Okayama 701-0193, Japan

Phone number: +81-86-462-1111, Fax number: +81-86-463-3508,

E-mail address: s-ito@mw.kawasaki-m.ac.jp

## **Disclosures:**

Financial disclosure statements have been obtained, and no conflicts of interest have been reported by the

authors or by any individuals in control of the content of this article.

# Use of Electrical or Magnetic Stimulation for Generating Hip Flexion Torque

## ABSTRACT

**Objective:** The purpose of this study was to investigate the most suitable site and method to effectively generate isometric hip flexion torque (torque value) using transcutaneous electrical or magnetic stimulation.

**Design:** Eleven healthy volunteers underwent torque value and pain degree measurements during magnetic stimulation of the iliopsoas using 3 coil placements. Following that, the peak torque values generated under 3 conditions of electrical stimulation of the sartorius, tensor fasciae latae, and rectus femoris, or that generated by magnetic stimulation of the iliopsoas were recorded at maximum tolerance intensity.

**Results:** No significant differences in torque values were observed among the 3 coil placements. Magnetic stimulation of a point below the inguinal ligament caused significantly more pain than the other points. Magnetic stimulation of the iliopsoas generated significantly higher torque values than electrical stimulation of the 2 hip flexor muscles together.

**Conclusions:** The hip joint was one of the most suitable regions for application of magnetic stimulation, as an alternative method to electrical stimulation.

**Key Words:** Magnetic Stimulation, Transcutaneous Electrical Stimulation, Torque, Pain

## INTRODUCTION

The presence of brain plasticity in adults has been of particular interest in recent neurological research. Many studies have shown that neuromuscular electrical stimulation of muscles was a useful treatment for motor paralysis caused by central nervous system damage. The Japanese Guidelines for the Management of Stroke (2009) have recommended electrical stimulation as an adjunct therapy with the usual rehabilitation exercises, as a result of much evidence.<sup>1</sup>

Transcutaneous functional electrical stimulation (FES) techniques applied for improving gait are roughly divided into 2 trials: single-channel and multi-channel stimulation. Trials using single-channel stimulation primarily focused on controlling the peripheral ankle joint.<sup>2, 3, 4</sup> Patients with severe hemiplegia, who had low muscle tone, are excluded from application of single-channel stimulation. On the other hand, multi-channel stimulation technique was applied for restoring patient's walking ability and demonstrated several outstanding effects.<sup>5-9</sup> However, this method was not clinically widespread because of the technical difficulty in the control of multiple joints using only electrical stimulation. In addition, the stimulation apparatus was very large and expensive for use in clinical settings and skilled techniques were required to operate the stimulus system. These factors have prevented the application of this method in clinical sites. Transcutaneous FES has also fatal limitation that this method cannot contract the iliopsoas (IL), the prime mover of hip joint flexion, because the IL is located too deep to be directly stimulated by surface electrodes. Normal persons walking at their preferred speed may display no significant flexor

muscle action after initiating the first step<sup>10</sup> while the patients with severe paralysis are likely to need more efforts to induce hip flexional motion because of the lack of pendulum movement in lower extremities. The IL which has the most extensive cross-sectional area in hip flexors is useful to induce hip flexion movement effectively.

Recently, some studies have reported the use of not only electrical stimulation but also magnetic stimulation as external stimulations for muscle contractions. The studies have described the application of magnetic stimulation of the lower extremities via the femoral nerve<sup>11</sup> or quadriceps femoris muscles<sup>12, 13</sup>; the knee extension torque was measured to investigate the effect of this new application. Although magnetic stimulation is minimally invasive and can induce inner muscle contraction, no reports have stated that it was useful for stimulating the IL, which generates hip flexion torque.

In the clinical gait training of severe hemiplegic patients, knee-ankle-foot orthoses (KAFO) are used to compensate for the loss of stability in the paralytic lower extremities, and therapists assist the swing of the paralyzed lower extremities using their own feet to compensate for the loss of voluntary movements. However, it is difficult for a therapist to precisely assist the swing of the paralyzed lower extremity during gait training because the amount of the therapist's assistance is sometimes excessive to keep a patient standing by him or herself. Circumduction gait with external rotation of the hip joint is a typical abnormal gait pattern for hemiplegic patients. External rotation of the hip joint is caused secondarily by posterior rotation of the pelvis in the stance phase and is thought to be a negative effect of motor learning.

Therapists must repeatedly provide normal movement patterns and avoid abnormal movement patterns as much as possible from the first exercise.

The present study provides fundamental research to assist the swing of paralyzed lower limbs and to model a normal swing pattern during gait training from the point of view that control of a proximal single joint using electrical or magnetic stimulation is practical. The purpose of this study was to determine the most suitable method to effectively generate hip flexion torque using external stimulation. Therefore, we first compared maximum isometric hip flexion torque (torque value) and the degree of pain in different coil placements for magnetic stimulation. Furthermore, we compared torque values generated by 3 electrical stimulations of the superficial hip flexor muscles with magnetic stimulation of the IL to determine the most suitable technique for hip flexion.

## **SUBJECTS AND METHODS**

### **Measurement method of torque values**

Eleven healthy young men with neither neurological nor orthopedic disabilities in their lower extremities and trunks participated in this study. The mean  $\pm$  standard deviation values for age, height, and weight were  $19.9 \pm 1.3$  years,  $168.3 \pm 3.9$  cm, and  $61.0 \pm 5.8$  kg, respectively. Before the study began, all participants were adequately explained the study's purpose and methods before participation, and each of them provided written informed consent. The study was approved by our institution's research ethics

committee for human subjects.

After the identification of the stimulus sites for magnetic and electrical stimulation in the supine position as described below, torque values of the right hip flexors were randomly measured thrice in each participant during external stimulation. The participants rested for 2.5 min between individual tests. An isokinetic dynamometer (BIODEX SYSTEM 3; Sakai Medical Co. Ltd., Japan) was used to measure the torque value in the standing position (Figure 1).<sup>14</sup> The truncal forward and backward moments were prevented using a monitor of BIODEX SYSTEM3 as the feedback method of torque waves during rest period. The participants were ordered not to contract the hip flexors voluntarily during external stimulation. The averages of 3 torque values acquired from individual measurements were analyzed.

#### **Determination of the most suitable site for magnetic stimulation**

To determine the most suitable sites on the IL for magnetic stimulation, 3 stimulus points of the IL were selected according to the needle electrode insertion sites used in clinical electromyography<sup>15</sup> and palpation placement (Figure 2).<sup>16</sup> Point (1) and point (2) were located by palpation, and their midpoint was considered as point (3). Magnetic stimulation was administered by a repetitive magnetic stimulator (MagPro; Medtronic Inc., USA). A round magnetic coil with a 10-mm inner radius and a 60-mm outer radius (DANTEC Medical Inc., Denmark) was used. To inhibit coil heating during measurements, the stimulation frequency was set at 25 Hz with an on-time of 2 sec and an off-time of 15 sec. Since peak eddy current was reported to flow through near the center of the coil in the manufacturer's instruction

book, the center of the coil was placed on the 3 stimulus points of the IL and stuck to the skin surface as closely as possible. The site of nerve excitation was reported to depend on the direction of the nerve fibers and the coil geometry.<sup>17, 18, 19</sup> Accordingly, we examined optimal directions of the coil to get strong reactions and not to disturb the torque measurements. After the maximum tolerable intensity was determined for the 3 coil placements by increasing the intensity in 15-A/ $\mu$ s intervals, the lowest of the 3 intensities was selected for measurement of torque values. As a result, the stimulation intensity was set at 60 A/ $\mu$ s for all participants. Three times of stimulations were delivered at each placement of the coil. In addition, the degree of pain during magnetic stimulation was evaluated using the Wong-Baker FACES pain rating scale (face scale)<sup>20</sup> after each measurement. Face 5 indicated “hurts as much as you can imagine,” whereas face 0 indicated “no hurt.” To confirm whether the femoral nerve was excited or not by magnetic stimulation, we tried to record compound muscle action potentials (CMAPs) from the sartorius (SA) and the rectus femoris (RF) as a preliminary experiment. In fact, the amplitudes of CMAPs were detected on the recording electrodes placed on these muscles especially during the stimulation of point (1). Consequently, to generate the highest hip flexion torque had priority over other things in the current study.

#### **Comparisons of electrical and magnetic stimulation**

After the adequate placement of the coil was determined, the torque values generated by electrical and magnetic stimulation were compared. Electrical stimulation was delivered using a stimulator (ES-510; Ito Co. Ltd., Japan), after 2 self-adhesive electrodes (5 × 9 cm) were placed at 3 different conditions (Figure

3).<sup>14</sup> The motor points of the SA, RF, and tensor fasciae latae (TF) were previously searched for using another stimulator (CX-3; OG Giken Co. Ltd., Japan) in the supine position to determine the most contractible sites by electrical stimulation.

The parameters of the external stimulation procedure were frequency, 30 Hz; on-time, 2 sec; and off-time, 15 sec, as described by Han et al.<sup>12</sup> and Szecsi et al.<sup>13</sup> The intensity of each stimulation was increased in a stepwise manner in 5-mA increments for electrical stimulation and 15-A/ $\mu$ s increments for magnetic stimulation until the participants could no longer tolerate the pain (maximum tolerable intensity). The stimulus site of the IL, at which the maximum torque value was produced in the first half of the present study, was adopted as a representative IL site to compare torque values between electrical and magnetic stimulation. Prior to the torque measurement during magnetic stimulation, the torque measurements during electrical stimulation were conducted. The stimulation sequence under the 3 electrical stimulus conditions was random.

## **Statistical Analysis**

SPSS 15.0J for Windows (SPSS Japan Inc., Japan) was used for statistical analysis. A one-way repeated-measures analysis of variance was used to compare torque values among the 3 coil placements and those between electrical and magnetic stimulation methods. The Friedman test was used to compare the degrees of pain experienced. The multiple comparison tests were performed when significant differences were found. Values of  $P < 0.05$  were considered statistically significant.



126

## 127 **RESULTS**

### 128 **Investigation of the most suitable site for magnetic stimulation**

129       The individual torque value data obtained with the 3 coil placements are presented in Table 1. The  
130       mean torque values for point (3) were the highest, followed by point (1) and point (2). Peak torque was  
131       induced in 5 participants each at point (1) and point (3) and in 1 participant at point (2). Thus, there were  
132       no significant differences in torque values among the 3 coil placements (Table 2). With regard to the  
133       degree of pain, we found that magnetic stimulation of point (1) caused significantly more pain than that at  
134       point (2). However, significant differences were not observed among other stimulus sites (Table 2). The  
135       maximum pain ratio among all participants was face 4 (“hurts a whole lot”).

### 136 **Comparisons of hip flexion torque generated by electrical and magnetic stimulation**

137       Point (3) was selected as the site for magnetic stimulation of the IL. The mean torque value and  
138       standard deviation of SA + TF, SA + RF, RF + TF, and IL were  $12.8 \pm 6.0$  Nm,  $10.8 \pm 4.4$  Nm,  $12.0 \pm 4.4$   
139       Nm, and  $19.2 \pm 8.8$  Nm, respectively (Table 3). Magnetic stimulation of the IL generated significantly  
140       higher torque values than electrical stimulation of the SA + RF, RF + TF ( $P < 0.01$ ), and SA + TF ( $P <$   
141        $0.05$ ), although the pain induced by magnetic stimulation was same degree as that induced by electrical  
142       stimulation. No significant differences were noted among other conditions.

143

## DISCUSSION

### Investigation of the most suitable site for magnetic stimulation

Magnetic stimulation is known to induce eddy currents in vivo using time-varying magnetic fields and to excite nerves and muscles without stimulating skin nociceptors.<sup>21, 22</sup> In the present study, magnetic stimulation was used to contract the IL, which was difficult to stimulate by transcutaneous FES. Although the peak torque was generated in 5 participants at point (1) or point (3), it was generated in only 1 participant at point (2). Thus, no significant differences in torque values were observed among the 3 different coil placements. The femoral nerve runs between the psoas and the iliacus muscles in the proximal part of the inguinal ligament and reaches the anterior part of the thigh through the muscular space. It branches off and innervates the psoas major and iliacus in the minor pelvis.<sup>23</sup> Because the motor point of the IL is located in the upper part of the inguinal ligament, it was anticipated that point (2) or point (3) were suitable sites for coil placement in the case of IL stimulation. However, the torque value at point (2) tended to be lower than that at the other stimulation sites. Contraction of the rectus abdominis seemed to be stronger than that of the IL by observation because point (2) was the nearest position to the rectus abdominis and, moreover, might be the farthest position from the IL due to structural feature of pelvis. The rectus abdominis should be suppressed to contract in order not to cause new gait disturbance by use of magnetic stimulation. These causes therefore seemed to indicate that point (2) was the unsuitable site of stimulation.

Regarding the degree of pain, stimulation of point (1) was more likely to induce pain than the other points. The pain factor caused by magnetic stimulation directly stimulated some nociceptors: A-delta myelinated heat nociceptors and C-fiber nociceptors in the muscle, tendon, and fascia. Han et al.<sup>12</sup> and Szecsi et al.<sup>13</sup> have reported that magnetic stimulation caused not only muscle contraction but also some degree of stimulation-induced pain. The stimulus intensity of the thigh muscles reported in previous studies was higher than that of the lower abdomen reported in this study. This indicated that pain sensitivity varied with the stimulation site and that the number of nociceptors affected the degree of pain during magnetic stimulation. Therefore, it is assumed that the number of nociceptors under the epidermis of point (1) was higher than that of other stimulus points.

The round coil used in this study had a diameter of 14 cm; therefore, it was difficult to exclude the influence of its stimulation on other sites. Future research involving mapping of the motor points of the IL should be performed using an 8-figure coil to investigate the best stimulation site.

#### **Application of magnetic stimulation**

Electrical stimulation of the quadriceps femoris muscle was reported to generate larger knee extension torque than magnetic stimulation in patients with a spinal cord injury and complete sensory loss.<sup>13</sup> However, the torque value generated during magnetic stimulation was larger than that generated during electrical stimulation in patients with partial sensory loss or without sensory disturbances.<sup>12, 13</sup> The participants in the present study were healthy and had no sensory problems, and hence,

stimulation-induced pain appeared to be a major factor restricting generating torque. The results of this study are consistent with those of previous studies, suggesting that magnetic stimulation is a low-invasive method<sup>21, 22</sup> even if the stimulus intensity is set at the maximum tolerance intensity of the individual subjects.

The advantage of electrical stimulation is that it can simultaneously stimulate plural muscles in the superficial layer, whereas magnetic stimulation can induce deep muscle contraction. In this study, magnetic stimulation of the IL generated larger torque values than electrical stimulation of the 2 hip flexor muscles in the superficial layer together. Our results suggested that the hip joint was one of the most suitable sites for magnetic stimulation as an alternative to electrical stimulation.

With regard to inducing the paralyzed lower extremity ahead during the swing phase, previous studies have reported that hip flexion increases when the action of plantar flexion decreases.<sup>24, 25</sup> During gait training of patients with severe hemiplegia, the ankle joint is usually controlled by ankle-foot orthosis (AFO) or by KAFO. Because of the weight of an orthosis and the compensation of plantar flexion torque, the hip flexion torque required in the early swing phase might be greater for patients using an orthosis compared to those not using it. The mean torque value generated in this study was  $19.2 \pm 8.8$  Nm. It may be inadequate to induce the lower extremity ahead because the KAFO weight and the abnormal muscle tone cause difficulty of affected hip flexion at the swing phase. The use of a combined method of electrical stimulation to the SA, TF, and RF, and magnetic stimulation should be considered in the future.

198        Impairment of patients with hemiplegia is much more severe in distal parts than in proximal parts.<sup>26</sup>  
199        Hip and plantar flexion greatly influence an individual's walking speed.<sup>27, 28</sup> The use of electrical and  
200        magnetic stimulation to hip flexors in patients with severe hemiplegia is anticipated to strengthen the  
201        weak hip flexors or to augment motor control aside from the application during gait training. For the  
202        purpose of clinical use, more trials to find the best spot for increasing the hip flexion torque and to  
203        decrease pain by moving the stimulation coil on each subject must be needed. Additionally, we should  
204        consider the kinematic and kinetic action of the hip flexors during gait and, moreover, investigate subjects  
205        and therapeutic protocols of magnetic stimulation in the future.

206

## 207        **REFERENCES**

- 208        1. Eto F, Dohi N, Ishigami S, Ito T, Kimura A, Tyuma T, et al: Rehabilitation, in Shinohara Y, Ogawa A,  
209        Suzuki N, Katayama Y, Kimura A (eds): Japanese Guidelines for the Management of Stroke 2009, 1st  
210        ed. Tokyo, Kyowa Kikaku, 2009, pp 272-338. Japanese
- 211        2. Liberson WT, Holmquest HJ, Scot D, Dow M: Functional electrotherapy: stimulation of the peroneal  
212        nerve synchronized with the swing phase of the gait of hemiplegic patients. Arch Phys Med Rehabil

213 1961; 42: 101-5

214 3. Burridge JH, Taylor PN, Hagan SA, Wood DE, Swain ID: The effects of common peroneal stimulation

215 on the effort and speed of walking: a randomized controlled trial with chronic hemiplegic patients. Clin

216 Rehabil 1997; 11: 201-10

217 4. Burridge JH, McLellan DL: Relation between abnormal patterns of muscle activation and response to

218 common peroneal nerve stimulation in hemiplegia. J Neurol Neurosurg Psychiatry 2000; 69: 353-61

219 5. Bogataj U, Gros N, Malezic M, Kelih B, Kljajic M, Acimovic R: Restoration of gait during two to three

220 weeks of therapy with multichannel electrical stimulation. Phys Ther 1989; 69: 319-27

221 6. Bogataj U, Gros N, Kljajic M, Acimovic R, Malezic M: The rehabilitation of gait in patients with

222 hemiplegia: a comparison between conventional therapy and multichannel functional electrical

223 stimulation therapy. Phys Ther 1995; 75: 490-502

224 7. Alon G, Ring H: Gait and hand function enhancement following training with a multi-segment

225 hybrid-orthosis stimulation system in stroke patients. J Stroke Cerebrovasc Dis 2003; 12: 209-16

- 226 8. Yan T, Hui-Chan CW, Li LS: Functional electrical stimulation improves motor recovery of the lower  
227 extremity and walking ability of subjects with first acute stroke: a randomized placebo-controlled trial.  
228 Stroke 2005; 36: 80-5
- 229 9. Malezic M, Hesse S, Schewe H, Mauritz KH: Restoration of standing, weight-shift and gait by  
230 multichannel electrical stimulation in hemiparetic patients. Int J Rehabil Res 1994; 17: 169-79
- 231 10. Perry J, Burnfield JM: Hip, in Perry J, Burnfield JM (eds): Gait analysis: normal and pathological  
232 function, 2nd ed. Thorofare, SLACK Incorporated, 2010, pp 103-19.
- 233 11. Kremenec IJ, Ben-Avi SS, Leonhardt D, McHugh MP: Transcutaneous magnetic stimulation of the  
234 quadriceps via the femoral nerve. Muscle Nerve 2004; 30: 379-81
- 235 12. Han TR, Shin HI, Kim IS: Magnetic stimulation of the quadriceps femoris muscle: comparison of  
236 pain with electrical stimulation. Am J Phys Med Rehabil 2006; 85: 593-9
- 237 13. Szecsi J, Schiller M, Straube A, Gerling D: A comparison of functional electrical and magnetic  
238 stimulation for propelled cycling of paretic patients. Arch Phys Med Rehabil 2009; 90: 564-70

- 239 14. Ito T, Tsubahara A, Seno Y, Tokuhiko H, Watanabe S: Consideration of ways to generate hip flexion  
240 torque by using electrical stimulation: Measurement of torque and the degree of pain. Jpn J Compr  
241 Rehabil Sci 2011; 2: 31-5
- 242 15. Perotto AO: Anatomical guide for the electromyographer: The limbs and trunk, 3rd ed. Niigata,  
243 Nishimura, 1997, pp 172-3. Japanese
- 244 16. Tixa S: Atlas of anatomy palpation: Hip joint, 2nd ed. Tokyo, Igaku-Shoin, 2007, pp 20-1. Japanese
- 245 17. Hsu KH, Nagarajan SS, Durand DM: Analysis of efficiency of magnetic stimulation. IEEE Trans  
246 Biomed Eng 2003; 50: 1276-85
- 247 18. Roth BJ, Basser PJ: A model of the stimulation of a nerve fiber by electromagnetic induction. IEEE  
248 Trans Biomed Eng 1990; 37: 588-97
- 249 19. Ueno S, Matsuda T, Hiwaki O: Estimation of structures of neural fibers in the human brain by  
250 vectorial magnetic stimulation. IEEE Trans Magn 1991; 27: 5387-920. Wong DL, Baker CM: Pain in  
251 children: comparison of assessment scales. Pediatr Nurs 1988; 14: 9-17



- 252 21. Baker AT, Freeston IL, Jalinous R, Jarratt JA: Magnetic stimulation of the human brain and peripheral  
253 nervous system. Neurosurgery 1987; 20: 100-9
- 254 22. Baker AT: An introduction to the basic principles of magnetic nerve stimulation. J Clin Neurophysiol  
255 1991; 8: 26-37
- 256 23. Kahle VW, Leonhardt W, Platzer W: Atlas of anatomy, 3rd ed. Tokyo, Bunkodo, 1995, pp431-3.  
257 Japanese
- 258 24. Lewis CL, Ferris DP: Walking with increased ankle push off decreases hip muscle moments. J  
259 Biomech 2008; 41: 2082-9
- 260 25. Nadeau S, Gravel D, Arsenault AB, Bourbonnais D: Plantarflexor weakness as a limiting factor of gait  
261 speed in stroke subjects and the compensating role of hip flexors. Clin Biomech 1999; 14: 125-35
- 262 26. Adams RW, Gandevia SC, Skuse NF: The distribution of muscle weakness in upper motoneuron  
263 lesions affecting the lower limb. Brain 1990; 113: 1459-76
- 264 27. Olney SJ, Griffin MP, Monga TN, McBride: Work and power in gait of stroke patients. Arch Phys

265 Med Rehabil 1991; 72: 309-14

266 28. Olney SJ, Griffin MP, McBride ID: Temporal, kinematic, and kinetic variables related to gait speed in  
267 subjects with hemiplegia: a regression approach. Phys Ther 1994; 74: 872-85

268

## 269 **FIGURE LEGENDS**

270 Figure 1. Measurement of torque value

271 The participants first stood on a 10-cm high platform, and the hip joint axis was matched to the machine's  
272 dynamometer axis. They were then told to stand half upright on their left leg. The distal part of the right  
273 thigh was fixed to the attachment with the right leg raised above the floor.

274 Figure 2. Stimulus sites on the iliopsoas for placement of magnetic stimulation coils

275 Point (1) was located at a distance of 2-fingers width lateral to the femoral artery (F. A.) and 1-finger  
276 width below the inguinal ligament (Ing. Lig.). Point (2) was located on the line connecting the navel with  
277 the anterior superior iliac spine (ASIS), beside the lateral site of the right rectus abdominis muscle. Point

278 (3) was the midpoint of point (1) and point (3).

279 Figure 3. Locations of the surface electrodes

280 The 2 electrodes were placed over individual motor points of 2 separate muscles. The following 3

281 conditions were selected for electrode placement. Conditions:

282 (1) The individual motor points of the sartorius and the tensor fasciae latae (SA + TF)

283 (2) The individual motor points of the sartorius and the rectus femoris (SA + RF)

284 (3) The individual motor points of the rectus femoris and the tensor fasciae latae (RF + TF)